## Red-D-Arc Weiderentals

## Calibration Certificate

Unit Serial Number: LK220086

Date Tested: Dec 20/2012

Instrument Number: 64592

Mac Address: 00:08:80:00:FC:50

IP Address:

Primary Standard						
	382 (F)		882 (F)		1382 (F)	
	Signal Voltage	Recorder display	Signal Voltage	Recorder display	Signal Voltage	Recorder display
Channel 1	2.784	387	6.00	980	9.236	1382
Channel 2	2.795	383	6.010	882	9,221	1380
Channel 3	2,798	384	6.004	881	9.206	1378
Channel 4	2.798	382	6.010	188	9,220	1380
Channel 5	2.785	382	6.025	884	9,221	1381
Channel 6	2.782	382	5.993	880	9,217	1380

Allowable difference in temp reading is +or- 5 Deg. F.

## **Instruments Used:**

Fluke 714 Thermocouple Calibrator SN: 2118058

Calibration Date: July 26, 2012

Precision digital multimeter

SN: 2142007

Dec 20,2013 A

Calibration Date: Awa 16, Zo12

ProHeat 35 power supply used for signal output SN:

Calibration date: NoV 30, 2012

MC 300 259 G

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.

Calibration due date: