

Red-D-Arc Welderrentals

Calibration Certificate

Unit Serial Number: MA040046G
 Instrument Number: 98802
 IP Address: 192.168.111.222

Date Tested: Nov 22, 2013
 Mac Address: 00:0A:8D:01:81:F2

Primary Standard						
	382 (F)		882 (F)		1382 (F)	
	Signal Voltage	Recorder display	Signal Voltage	Recorder display	Signal Voltage	Recorder display
Channel 1	2.788	382	6.013	882	9.239	1382
Channel 2	2.788	382	6.015	882	9.233	1382
Channel 3	2.778	381	6.014	882	9.226	1381
Channel 4	2.781	381	6.018	883	9.235	1381
Channel 5	2.782	381	6.012	882	9.240	1382
Channel 6	2.782	381	6.008	882	9.235	1382

Allowable difference in temp reading is +or- 5 Deg. F.

Instruments Used:

Fluke 714 Thermocouple Calibrator SN: 2118055

Calibration Date: 08/09/13

Precision digital multimeter SN: 2142011

Calibration Date: 08/09/13

ProHeat 35 power supply used for signal output SN: MC430044G

Calibration date: 11/22/13

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.

Calibration due date: 11/22/14

Technician: *David Burger*
 Signature