

Calibration Certificate

Unit Serial Number: MA120017G

Instrument Number: 95106 Mac Address: 00:0A:8D:01:73:82

Date Tested: 4/2/2013

IP Address: 192.168.111.222

| Primary Standard | | | | | | | |
|------------------|-------------------|---------------------|-------------------|---------------------|-------------------|---------------------|--|
| | 382 (F) | | 882 (F) | | 1382 (F) | | |
| | Signal Voltage | Recorder display | Signal Voltage | Recorder display | Signal Voltage | Recorder display | |
| Channel 1 | 2.785 | 382 | 6.009 | 881 | 9.227 | 1380 | |
| Channel 2 | 2.784 | 383 | 6.015 | 882 | 9.225 | 1380 | |
| Channel 3 | 2.790 | 382 | 6.013 | 882 | 9.223 | 1380 | |
| Channel 4 | 2.790 | 382 | 6.010 | 881 | 9.226 | 1380 | |
| Channel 5 | 2.790 | 382 | 6.012 | 882 | 9.225 | 1380 | |
| Channel 6 | 2.789 | 382 | 6.011 | 881 | 9.226 | 1380 | |

Allowable difference in temp reading is +or- 5 Deg. F.

Instruments Used:

| Fluke 714 Thermocouple Calibrator SN: N/A | Calibration Date: N/A |
|--|------------------------------|
| Precision digital multimeter SN: 2036021 | Calibration Date: 12/17/2013 |
| ProHeat 35 power supply used for signal output SN: N/A | Calibration date: N/A |

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

| Unit passed inspection. | Technician: Scott T. |
|--------------------------------|----------------------|
| Calibration due date: 4/2/2014 | Signature |