

Red-D-Arc Welderrentals

Calibration Certificate

Unit Serial Number: MB430148G
 Instrument Number: 128509
 IP Address: 192.168.111.222

Date Tested: 03/05/13
 Mac Address: 00:0A:8D:01:F5:FD

Primary Standard						
	382 (F)		882 (F)		1382 (F)	
	Signal Voltage	Recorder display	Signal Voltage	Recorder display	Signal Voltage	Recorder display
Channel 1	382	384	882	883	1382	1384
Channel 2	382	385	882	884	1382	1384
Channel 3	382	384	882	882	1382	1384
Channel 4	382	384	882	883	1382	1384
Channel 5	382	383	882	883	1382	1383
Channel 6	382	383	882	883	1382	1384

Allowable difference in temp reading is +or- 5 Deg. F.

Instruments Used:

Fluke 714 Thermocouple Calibrator SN: 2118055

Calibration Date: 26/07/12

Precision digital multimeter SN: 2142011

Calibration Date: 16/08/12

ProHeat 35 power supply used for signal output SN: MC430140G

Calibration date: 03/05/13

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.

Calibration due date: 03/05/14

Technician: Gus Roch

Signature

Montreal Canada office