

Red-D-Arc Welderrentals

Calibration Certificate

Miller ProHeat 35

Unit Serial Number: MC450063G

Date Tested: 06/16/15

Certified Calibration

Primary Standard						
	382° (F) 2.787VDC@RC9		882° (F) 6.013VDC@RC9		1382° (F) 9.239VDC@RC9	
	RC9 output DC voltage reading	ProHeat 35 display	RC9 output DC voltage reading	ProHeat 35 display	RC9 output DC voltage reading	ProHeat 35 display
TC1	2.790	382	6.015	882	9.238	1382
TC2	2.788	383	6.013	883	9.238	1382
TC3	2.795	383	6.017	883	9.243	1383
TC4	2.794	384	6.015	883	9.245	1383
TC5	2.794	383	6.021	883	9.244	1383
TC6	2.801	384	6.024	883	9.248	1383
Allowable	2.749-2.825	+/- 3° F	5.975-6.051	+/- 3° F	9.201-9.277	+/- 3° F

Instruments Used:

Fluke 714 Thermocouple Calibrator SN: 2118055

Calibration Date: 24/07/14

Precision Digital Voltage Meter SN: 2142011

Calibration Date: 24/07/14

Machine re-calibration due date: 06/16/16

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.



Technician Gus Rodin Date: 06-16-15

Signature